



महाराष्ट्रशासन
SIR J. J. GROUP OF HOSPITALS, BYCULLA,
MUMBAI - 400 008
Medical Store
सर ज.जी. समुह रुग्णालये, भायखला, मुंबई-४००००८
Tel.No. ०२२२३७३९१०३, ०२२२३२२४५३, ०२२२३७३५५५५
Email id-jjhmedicalstore@gmail.com



No. JJH/SS/QTN/ 1246 /2025

Date :- 30 / 07 /2025

Sub : Quotation for the supply of Surgical /items (Open Quotation)

Sir,

You are requested to submit your lowest bid for Surgical / items. The quotation should reach this office in a sealed envelope on or before 07 / 08 / 2025 till 5.00 pm. Quotation is also published on www.ggmcijh.com

The Dean, Sir J J Group of Hospitals, Mumbai reserves the right to Accept ,Recall or Reject any or all quotations without assigning any reason. Other instructions and terms & conditions regarding quotation are mentioned below the Surgical items list.

Sr. No.	Name of Item	Specification	Pack Size	Rate per pack size (Inc. GST)	MRP
1	Sterile Disposable Needle Size - 18 G	Single use Sterile, surface should be extremely smooth,the needle point is extremely sharp.Color coded, individually blister packed, hypodermic needle 1 Inch	Each		
2	Sterile Disposable Needle Size -20 G	Single use Sterile, surface should be extremely smooth,the needle point is extremely sharp.Color coded, individually blister packed, hypodermic needle 1 Inch	Each		
3	Sterile Disposable Needle Size - 21 G	Single use Sterile, surface should be extremely smooth,the needle point is extremely sharp.Color coded, individually blister packed, hypodermic needle 1 Inch	Each		
4	Sterile Disposable Needle Size - 22 G	Single use Sterile, surface should be extremely smooth,the needle point is extremely sharp.Color coded, individually blister packed, hypodermic needle 1 Inch	Each		
5	Sterile Disposable Needle Size - 26 1/2 G	Single use Sterile, surface should be extremely smooth,the needle point is extremely sharp.Color coded, individually blister packed, hypodermic needle 1/2 Inch	Each		
6	Sterile Scalp Vein Set (Disposable) Size 18G	Sterile Scalp Vein Set (Disposable) Size 18G	Each		
7	Sterile Scalp Vein Set (Disposable) Size 21G	Sterile Scalp Vein Set (Disposable) Size 21G	Each		
8	Sterile Scalp Vein Set (Disposable) Size 22G	Sterile Scalp Vein Set (Disposable) Size 22G	Each		
9	Sterile Scalp Vein Set (Disposable) Size 23G	Sterile Scalp Vein Set (Disposable) Size 23G	Each		
10	Sterile Scalp Vein Set (Disposable) Size 24G	Sterile Scalp Vein Set (Disposable) Size 24G	Each		
11	Infant Feeding Tube Size -9 FG	Infant Feeding Tube Size -9 FG	Each		
12	Sterile Suction Catheter- 10 FG	Sterile Suction Catheter- 10 FG	Each		
13	Sterile Suction Catheter- 12 FG	Sterile Suction Catheter- 12 FG	Each		
14	Sterile Suction Catheter- 14 FG	Sterile Suction Catheter- 14 FG	Each		
15	Sterile Suction Catheter- 16 FG	Sterile Suction Catheter- 16 FG	Each		
16	Steril Disp. Umbilical Cord Clamp	Steril Disp. Umbilical Cord Clamp	Each		
17	Airway Plastic Latex free- 2	Airway Plastic Soft Plastic Nasal Airway are constructed of a soft latex free plastic. The interior of the airway has striations for easy suctioning and insertion. Latex free size 2	Each		
18	Airway Plastic Latex free- 3	Airway Plastic Soft Plastic Nasal Airway are constructed of a soft latex free plastic. The interior of the airway has striations for easy suctioning and insertion. Latex free size 3	Each		



	Name of Item	Specification	Pack Size	Rate per pack size (Inc. GST)	MRP
19	Airway Plastic Latex free-4	Airway Plastic Soft Plastic Nasal Airway are constructed of a soft latex free plastic. The interior of the airway has striations for easy suctioning and insertion. Latex free size 4	Each		
20	Airway Plastic Latex free- 5	Airway Plastic Soft Plastic Nasal Airway are constructed of a soft latex free plastic. The interior of the airway has striations for easy suctioning and insertion. Latex free size 5	Each		
21	Triple Lumen Catheter 7 Fr	1. Should be Radio-opaque catheter material made of polyurethane with specially designed soft tip 2. Kink resistant guide wire with soft J-tip 3. Guidewire should be made of nitinol or equivalent 4. length 15/20cm, 18G valve needle having side port 5. Safety connector	Each		
22	Plain Endotracheal Tube No 2	Endotracheal Tube Cuffed PVC Transparent Sterile Disposable for pediatric use size-2 mm ID	Each		
23	Plain Endotracheal Tube No 2.5	Endotracheal Tube Cuffed PVC Transparent Sterile Disposable for pediatric use size-2.5 mm ID	Each		
24	Plain Endotracheal Tube No 3	Endotracheal Tube Cuffed PVC Transparent Sterile Disposable for pediatric use size- 3 mm ID	Each		
25	Plain Endotracheal Tube No 3.5	Endotracheal Tube Cuffed PVC Transparent Sterile Disposable for pediatric use size-3.5 mm ID	Each		
26	Plain Endotracheal Tube No 4	Endotracheal Tube Cuffed PVC Transparent Sterile Disposable for pediatric use size-4 mm ID	Each		
27	Cuff Endotracheal Tube No 6.5	Endotracheal Tube Cuffed PVC Transparent Sterile Disposable for adult use size-6.5 mm ID	Each		
28	Cuff Endotracheal Tube No 7	Endotracheal Tube Cuffed PVC Transparent Sterile Disposable for adult use size- 7 mm ID	Each		
29	Cuff Endotracheal Tube No 7.5	Endotracheal Tube Cuffed PVC Transparent Sterile Disposable for adult use size-7.5 mm ID	Each		
30	Cuff Endotracheal Tube No 8	Endotracheal Tube Cuffed PVC Transparent Sterile Disposable for adult use size- 8 mm ID	Each		
31	Cuff Endotracheal Tube No 8.5	Endotracheal Tube Cuffed PVC Transparent Sterile Disposable for adult use size-8.5 mm ID	Each		
32	Ryles Tube No 12	Ryles tube Sterile Disposable size-12fr	Each		
33	Incentive Spirometer	Should have three color-coded balls in each chamber to provide a visual incentive for patients. Air should enter each chamber and raise the balls depending on the flow of air inhaled per second. Available in lung capacity : 600ml/sec, 900 ml/sec and 1,200ml/sec. The whole set should include EVA tubing, mouth piece 13M, tri-ball Incentive spirometer. Single-patient use to minimize the risk of cross contamination. 12 mm OD connector Volume should be 600 mL/sec, 900 mL/sec, 1200 mL/sec Material: Should be AS for chamber, PE for ball and mouthpiece, and EVA for tubing	Each		
34	Bivalve (Threeway)	1. The stopcock should be manufactured from medical grade, clear transparent lipid compliant polycarbonate 2. Its three arm handle should be easy to operate 3. The device should allow for minimum residual volume 4. There should be a rigid stopcock with circular flow-channel prevents the incidence of air-trap 5. It should be provided with two female luer lock ports and one male luer lock port with rotating lock, for safe connection to other device 6. It should be sterile and individually packed in blister packplastic with luer lock & side connection	Each		

Sr. No.	Name of Item	Specification	Pack Size	Rate per pack size (Inc. GST)	MRP
35	Dial Flow	<p>The flow regulator should comprise of a collared long Y-site with a self-sealing latex-free septum which provides an optimal secondary site for bolus IV Injection</p> <p>The device comprises of a specially designed male luer with ribs which allows for connection to patient-side IV access and female luer with wings for easy gripping and attachment to IV fluids access.</p> <p>The regulator should have an ergonomic dial design with ridges that allows adjustment to control the flow rates from 20 to 250 ml/hr and double scale for 10% and 40% solution strengths</p> <p>The product should have a drip changer made from DEHP free material with an in-built 15-micron disc filter meant for particle filtration. The device should be non-pyrogenic and disposable but could be sterilized by Ethylene Oxide (ETO)</p>	Each		
36	Nasal Cannula (Rams Cannula)	Should be suitable for easy and efficient administration of oxygen. Should have kink resistant main tube to avoid accidental blockage, soft twin prong nasal tips to ensure equal distribution of oxygen through both air passages which provides maximum patient comfort.	Each		
37	Eye Drape	Set included with integrated fluid collection pouch in drape, size 170x260 cm, Made of three layers – Non Woven (hydrophilic) + PE Film Impermeable + Poly propylene (Nonwoven), for highest protection at the time of procedure or Equivalent Exceed EN 13795 US FDA certified / CE Marked	Each		
38	Ophthalmic Tunnel Blade (Crescent/Keratome/15 Slit/MVR)	<p>1) 1 mm MVR blade with handle is required for the side port incision. It should be bibeveled, spear shape 20 gauge, ophthalmic micro surgical full handled knife.</p> <p>2) 3.2 mm Keratome or slit knife with handle is required for the entry in to anterior chamber, It should be pointed tip angled Bevel up, ophthalmic microscope full handled knife.</p> <p>3) 2.6mm Crescent Blade is required for dissecting tunnel. It should be angled bevel up, ophthalmic microsurgical full handled knife. All blades should be etio-sterilized and disposable. All handles should be autoclavable</p> <p>For one Surgery a set of three blades is required.</p>	Each		
39	Trolley Drape	Should be disposable, sterile drape made from nonwoven fabric separately packed in EO sterile pouch	Each		
40	Laryngoscope Adult	Should be LED laryngoscope set, adult, macintosh type, Stainless steel blades with removable led lamp. Handle and blades are made of 304 grade stainless steel. Dull/matt finish. Supplied with 4 blades size 1, 2, 3 and 4 and two batteries for handle. Should provide with 1 spare led lamp and batteries	Each		
41	Laryngoscope Paediatric	Should be LED laryngoscope set, paediatric, macintosh type, Stainless steel blades with removable led lamp. Handle and blades are made of 304 grade stainless steel. Dull/matt finish. Supplied with 4 blades size 1, 2, 3 and 4 and two batteries for handle. Should provide with 1 spare led lamp and batteries	Each		
42	Venous Extension line with 3way size-100cm	<p>Small bore, high pressure low volume extension tube, with male luer lock at one end and female luer lock at other end.</p> <p>Made up of Medical grade P.V.C</p> <p>Latex Free.</p> <p>BP/ISO/ISI /CE STANDARDS 100cm</p>	Each		

No.	Name of Item	Specification	Pack Size	Rate per pack size (Inc. GST)	MRP
43	Venous Extension line with 3way size- 150cm	Small bore , high pressure low volume extension tube , with male luer lock at one end and female luer lock at other end . Made up of Medical grade P.V.C Latex Free. BP/ISO/ISI /CE STANDARDS 150cm	Each		
44	Venous Extension line with 3way size- 200cm	Small bore , high pressure low volume extension tube , with male luer lock at one end and female luer lock at other end . Made up of Medical grade P.V.C Latex Free. BP/ISO/ISI /CE STANDARDS 200cm	Each		
45	Venous Extension line with 3way size- 50cm	Small bore , high pressure low volume extension tube , with male luer lock at one end and female luer lock at other end . Made up of Medical grade P.V.C Latex Free. BP/ISO/ISI /CE STANDARDS 50cm	Each		

Instructions and terms & condition regarding Quotation:-

- Interested vendors should submit the quotation as per given format only. Vendors need to specify Manufacturer & brand in the quotation and the same supply needs to be supplied.
- Quotations must be submitted in a sealed envelope only.
- The quotation & envelope should be addressed to - The Dean, Sir J J Group of Hospitals, Mumbai and marked **Kind attention to- Surgical store.**
- Vendors must write quotation reference no & Last date of submission of quotation on the envelope.
- Any amendments regarding the quotation will be published on website www.ggmcjjh.in. Vendors will not be communicated separately regarding the amendments.
- However if the vendor fails to check any of these amendments on the website then it will be presumed that the vendor has quoted his/ her rates by taking the note of these amendments.
- Rate should be quoted **inclusive of all taxes ,GST , etc.**
- Rate must be written in both figures & words. Rates should be valid for six months from the date of opening the quotation.
- Rate must be quoted for Surgical Items & Instrument complying with the standards for Surgical Items & Instruments (ISO/ISI, etc. Whichever is applicable)
- Delivery period is 24 hrs to 03 (Three) days from the receipt of order, as per vitality of Surgical items.
- Analysis test reports (**Inhouse and NABL**) need to be submitted at the time of goods supply,
- For goods supplied under MJPJAY Other schemes , bills will be passed only after the utilization certificate is received.
- Successful vendors if fails to supply the goods within stipulated delivery period he is liable for the further necessary action, which may deem fit.
- Frequent defaulter bidder will be debarred from participating in the future tenders / quotations called by this office
- Sample needs to be submitted whenever asked.
- Following documents need to be submitted alongwith quotation- (To be submitted in Separate Envelope)**
 - Bidders FDA/Medical Device license (manufacturing license OR wholesale stocking & Selling Licence)
 - Manufacturers Authorization Letter is mandatory for all above mentioned products.**
 - Demo Sample will be require to submit to Surgical Store, Sir J J Group of Hospitals, Mumbai for all quoted items. The submitted material will not be return back to the vendor irrespective of qualification or disqualification
 - No conviction certificate from FDA
 - GST registration copy
 - Bidder details as per Annexure A
 - Adata (अदाता) Registration Number at JJ Hospital Mumbai
 - WHO GMP certificate of manufacturer, ISO/ISI/CE/EN/ASTM Product Standard Certificate whichever is applicable

NOTE : Last date of submission of quotation: 07 /08/2025 before 5.00 pm


Dean
Sir J J Group of Hospitals,
Mumbai

4/5



Annexure – A

(To be submitted on Bidder's Letterhead, Incomplete Annexure is liable for Rejection)

10. Name and address of the firm: -
11. Registered Head Office Postal address: -
12. Telephone No.
13. E-Mail ID : -
14. Ownership status of the firm- (Maharashtra Govt. / Central Govt./It. Sector /co - operative /SSI /Private)
15. Whether bidding as a manufacturer / Importer / Authorized Distributor
16. Name of the person & Phone no. who should be contacted by this office in case of emergency.
17. Payee (अदाता) Registration Number at Sir J J Hospital, Mumbai.
18. Bank Details: -
 - 1) Bank A/C No. _____
 - 2) IFSC Code: - _____
 - 3) Branch Name & Address: _____
 - 4) Cancelled Cheque: _____
10. PAN number _____
11. GST registration number _____

I / we hereby declare that particulars furnished above are true to the best of my /our knowledge and belief and that if any of the particulars is found to be materially incorrect / misleading, My /Our quotation shall be rejected. I / we accept all term & condition, also I / we are liable for penal action as per terms specified in the " terms and conditions of quotation".

Date: -

Signature of the bidder with official seal and address