



महाराष्ट्र शासन
SIR J. J. GROUP OF HOSPITALS, BYCULLA,
MUMBAI - 400 008
Surgical Store/ Medical Store
सर ज.जी. समुह रुग्णालये, भायखला, मुंबई-४००००८
Tel.No.०२२२३७३९१०३, ०२२२३२२२४५३, ०२२२३७३५५५५
Email id-jjhmedicalstore@gmail.com



No. JJH/SS/QTN/ 133 /2026

Date : 21/10/2026

Sub : Quotation for the supply of Surgical items (Open Quotation)

Sir,

You are requested to submit your lowest bid for Surgical / items. The quotation should reach this office in a sealed envelope on or before 09/11/2026 till 5.00 pm. Quotation is also published on www.ggmccjjh.com

The Dean, Sir J J Group of Hospitals, Mumbai reserves the right to Accept, Recall or Reject any or all quotations without assigning any reason. Other instructions and terms & conditions regarding quotation are mentioned below the Surgical items list.

Sr No	Name of Items	Specification	Packing	Rates per pc (inc. GST)	MRP
1	Anatomical Mask	an inflatable cuff for comfort, a standard 22 mm connector, and a disposable or autoclavable (up to 137°C) design, typically made from non-toxic silicone or PVC. 0,1,2,3,4			
2	CPAP Mask	1. The mask should have headgear with a provision to support on the patient face and reduce excess pressure on the patient's nose bridge. 2. Should have provision to maintain 12mm gap between forehead and mask nose bridge frame. 3. The headgear should have an option of providing a personalized fit to help increase comfort and tolerance for patients. 4. Headgear should help in easy sizing fitting and readjusting. 5. Mask should have auto adjusting seal and come with a sizing guide to choose right size for patients. 6. Should have provision for slot for Oro-Naso-Gastric Tubes.			
3	Ultrasonography Jelly	Should have neutral pH, be easy to spread, formaldehyde and salt free, nongreasy, proper adhesion for electrodes, easy squeeze leakproof bottle 250ml quantity, hypoallergenic, nontoxic odourless, high viscosity, high conductivity, water soluble, non staining easy to clean, non irritant to skin. Propylene glycol, Glycerine and phenoxyethanol containing gel Volume 250ml in compressible dispenser			
4	ECG Jelly	Should have neutral pH, be easy to spread, formaldehyde and salt free, nongreasy, proper adhesion for electrodes, easy squeeze leakproof bottle 250ml quantity, hypoallergenic, nontoxic odourless, high viscosity, high conductivity, water soluble, non staining easy to clean, non irritant to skin. Propylene glycol, Glycerine and phenoxyethanol containing gel Volume 250ml in compressible dispenser			
5	Endotracheal Tube Plain (2-9)	Should be available as a transparent tube with radiopaque line in extra-soft PVC. The distal end of the tube should be bevelled at 37+ - 10. The tube should have clear distance markings A standard male ISO adaptor 15 mm to connect the endotracheal tube to the ventilator The distal end of the tube should be bevelled at 37+/-10 2 to 9 mm ID			
6	Endotracheal Tube Cuffed (2-9)	Should be Transparent with Blue radio-opaque line. Should have Low pressure high volume cuff, reduces risk of creasing on inflation thus minimises pressure on tracheal wall. Printed pilot balloon, Smooth Murphy Eye. Should be able to verify tube position within patient. Misting can be clearly seen to confirm intubation The distal end of the tube should be bevelled at 37+/-10 2 to 9 mm ID			
7	Tracheostomy Tube cuffed (Size 6 - 8.5)	1. The Tracheostomy Tube should be a cuffed tube made of thermosensitive material which softens at body temperature to accommodate individual patient anatomy 2. It should have a soft flexible flange which helps in easy fixation 3. It should be provided with a 15 mm connector and can be rotated in 360 degree direction and facilitate the safe connection to breathing system from any direction 4. It should have a radio opaque blue line allows for clear			

Sr No	Name of Items	Specification	Packing	Rates per pc (inc. GST)	MRP
		X-ray visualization 5. It should be supplied with cotton neck tape and is sterile and individually blister packed 6. It should be light in weight and non irritant 7. The soft low pressure cuff should dynamically adjust the pressure at the point of contact during the breathing cycle. 8. It should have a pilot balloon with non return valve which facilitates the cuff inflation & deflation size- 6-8.5 mm			
8	Extension Line 25cm	Small bore , high pressure low volume extension tube , with male luer lock at one end and female luer lock at other end . Made up of Medical grade P.V.C, Latex Free. BP/ISO/ISI /CE STANDARDS 150cm			
9	Gigli Saw Wire	Made of martensitic steel (quenched, magnetic steel) and not stainless steel.Hooks at both ends.Should conform to ISO standard			
10	IV Cannula (Size 16 G)	Disposable, Sterile, Single use, individually transparent packing. Colour code according to size. Should have injection port, Snap lid on port, Leak proof, Must be leur fitting Adaptable fixation wings. Material polytetrafluoroethylene (PTFE)/ FEP. Should be non thrombogenic, Non Toxic and Non Pyrogenic. The inner needle must have a bevel for minimal puncture trauma Blood flow through needle must be adequate to ensure immediate flow Cannula Should be : a. Thin walled b. X - ray contrast line c. Smooth surface BP/ISO/ISI /CE STANDARDS 16g			
11	IV Cannula (Size 18-22G)	Disposable, Sterile, Single use, individually transparent packing Colour code according to size Should have injection port, Snap lid on port, Leak proof, Must be leur fitting Adaptable fixation wings. Material polytetrafluoroethylene (PTFE)/ FEP. Should be non thrombogenic, Non Toxic and Non Pyrogenic. The inner needle must have a bevel for minimal puncture trauma Blood flow through needle must be adequate to ensure immediate flow Cannula Should be : a. Thin walled b. X - ray contrast line, The distal end of the tube should be bevelled at 37+-10 Smooth surface BP/ISO/ISI /CE STANDARDS (Size 18G-22G)			
12	IV Cannula (Size 24 G)	Disposable, Sterile, Single use, individually transparent packing. Colour code according to size. Should have injection port, Snap lid on port, Leak proof, Must be leur fitting Adaptable fixation wings. Material polytetrafluoroethylene (PTFE)/ FEP. ææ Should be non thrombogenic, Non Toxic and Non Pyrogenic. The inner needle must have a bevel for minimal puncture trauma Blood flow through needle must be adequate to ensure immediate flow Cannula Should be : a. Thin walled b. X - ray contrast line The distal end of the tube should be bevelled at 37+-10 c. Smooth surface BP/ISO/ISI /CE STANDARDS 24G			
13	Ambu Bag Neonatal	It should be self-inflatable and must have pop up valve (non-return valve), attachment for oxygen tube & oxygen reservoir. Bag must be made up of medical grade silicon, latex free double layered which retain sensitivity and it must be resistant to rough use. Inlet end of the bag must have separate port for Oxygen supplement. Outer port must be such that re-breathing valve or non-return valve can be attached. Must be supplied with Oxygen reservoir bag. Should have a bag volume of 250ml. It should be autoclaveable. It should be adaptable to all type of face masks. Must be USFDA or European CE approved product			
14	Rubber Makintosh	Should be made of rubber with double colour, high grade polymer (55% grade A); Should be resistant to high temperature,chemical resistance and fair abrasion resistance.Should be water proof. Colour :Green,Blue, Red, Thickness-0.4mm, Width-90cm Roll of 30 mtr I.S.O.certified, Double colour,water proof ISI Mark 4135: 1974			
15	Mucus Extractor	Transparent Tube mucus extractor for neonatal & infants.Sterile Pyrogen Free,Non Toxic The capacity of the Mucus Extractor is 25ml and a spare cap is also provided for the sealing.Soft and kink resistant tubing with frosted surface A traumatic, soft rounded, beveled open tip. Capacity: 25 ml.			
16	Plastic Gloves	Disposable plastic gloves individually packed. Made of medical grade virgin P.V.C. not less than 50 micron thickness,non toxic with excellent grip Length: 7 Inch Free Size Sleeve Length: 295 mm mid-forearm Type: Non Sterile 100 pcs/Pack			

Sr No	Name of Items	Specification	Packing	Rates per pc (inc. GST)	MRP
17	Skin Grafting Blade	Stainless steel blade, sharp blade edge should offer smooth shaving of healthy skin tissue, Individually sterile packed, european CE certified, Expiry date should be written.			
18	T-Connector	T-Connector for Nebulization set			
19	T Piece	For use in recovery for the delivery of oxygen. Connects to 15 mm ET tube connector/ laryngeal mask. Fitted with 22 mm male / 15 mm female patient port. 6" flexible corrugated tube oxygen reservoir with oxygen supply port along with cap. Sterile and individually packed			
20	Flexometallic Endotracheal Tube (Size 6.5 to 8.5)	Used to maintain a patient airway and mechanically ventilate patients. Made from non-toxic, non-irritant medical grade pvc. Kink resistance. Tube can be temporarily straightend for ease of suctioning. Smooth polished murphy eye, distal end and tapered balloon end for a traumatic intubation and extubation Size 6.5 to 8.5 mmID			
21	Disposable Spinal Needle (Size 20 G)	Disposable, Sterile, Single use, Grooved hub with smooth movement, Length 90mm, Individually transparent packing. Color coded as per size, With introducer. BP/ISO/ISI /CE STANDARDS, 20G			
22	Disposable Spinal Needle (Size21G)	Disposable, Sterile, Single use, Grooved hub with smooth movement, Length 90mm, Individually transparent packing. Color coded as per size, With introducer. BP/ISO/ISI /CE STANDARDS, 21G			
23	Disposable Spinal Needle (Size22G)	Disposable, Sterile, Single use, Grooved hub with smooth movement, Length 90mm, Individually transparent packing. Color coded as per size, With introducer. BP/ISO/ISI /CE STANDARDS, 22G			
24	Disposable Spinal Needle (Size23G)	Disposable, Sterile, Single use, Grooved hub with smooth movement, Length 90mm, Individually transparent packing. Color coded as per size, With introducer. BP/ISO/ISI /CE STANDARDS, 23G			
25	Disposable Spinal Needle (Size24G)	Disposable, Sterile, Single use, Grooved hub with smooth movement, Length 90mm, Individually transparent packing. Color coded as per size, With introducer. BP/ISO/ISI /CE STANDARDS, 24G			
26	Disposable Spinal Needle (Size25G)	Disposable, Sterile, Single use, Grooved hub with smooth movement, Length 90mm, Individually transparent packing. Color coded as per size, With introducer. BP/ISO/ISI /CE STANDARDS, 25G			
27	Negative Wound Suction Set (Romovac Drain) Size 10-18G	Negative Suction Wound Drainage vacuum set should have graduated, bellow of 600 ml capacity, should have 2 radio opaque lined drain catheter Size-10-18G respectively for No.10-18G Should be double packed sterile			
28	N-95 Mask	Should have a shape that should not collapse easily. High filtration efficiency, Good breathability, Tight fit. Quality compliant with standards of surgical N-95 Mask, Fluid resistance at minimum 80mmHg pressure based on ASTM F1862, Quality compliant with standard for particulate respirator that can be worn with full face shield. They should be tested for fluid resistance, filtration efficiency (particulate filtration efficiency and bacterial filtration efficiency), flammability and biocompatibility. NIOSH certified, EN 149:2001 European standard, ISO 22609 standard or equivalent.			
29	Triple Layer Face Mask	Should be triple layer face mask with rust proof nose clip Tie on Mask of Non-woven, hypoallergenic 3 ply construction with filter in between offering > 99% bacterial filtration efficiency (BFE) at 5 micron capacity. Standard size with 4 tie strings and elastic strings			
30	Pulse Oximeter	Pulse Oximeter			
31	Fistula Needle Set 17G	Should be ultra-siliconized thin walled sharp needle with beveled back eye for easy insertion and minimize trauma to the patient. Should be soft and kink resistant medical grade PVC tubing, available with color coded fixed wings, flexible butterfly wings for better grip, needle with back-eye, non toxic, non pyrogenic, sterile. Needle length - 30cms			
32	Sterile Plastic Drape	Sterile disposable poly plain Sheet 100cm x 150cm			
33	Sterile Surgical Gown	SFSMS Polypropylene nonwoven (Non laminated) breathable fabric with antistatic and alcohol repellent treatment, Raglan Sleeves, Adjustable Hook & Loop Binding, Flourine Chemical Free. CE / ISO			

Sr No	Name of Items	Specification	Packing	Rates per pc (inc. GST)	MRP
34	Shoe Cover	Made of Nonwoven material which is water and alcohol repellent. The pair should be of standard size, should be long upto below knee.			
35	Vacuum Blood Collection Tube- EDTA	Lavender double cap with spray dried 5.4mg K2EDTA 13X75mm 3ml draw volume evacuated tube compatible and listed in approved tube list of automated hematology machines kit insert if applicable. Safe for Incineration: Environmental safety data must be provided. Material: tube PET (Poly ethylene terephthalate- virgin material). Holding capacity : 3ml or 4ml. Use: for blood Specimen collection. Standard Dimensions in Millimeters: Vacuum tube 13 mm (D) x 75mm (L). Error margin of biggest vacuum should be 10%, Shelf Life Not less than 12 months, Sterilization Sterile (Gamma radiation), Single Use. CE / ISO			
36	Yankauer Suction Set	Suction Cannula with Suction Tubing			
37	Sterile Cord Clamp	Disposable Umbilical Cord Clamp are made from Non toxic Medical grade polymer. 1. Material: PVC, PE, PP, ABS etc. 2. Sterilize by ethylene oxide gas 3. Supplied sterile in individual peel-able polybag. Or Equivalent			

Instructions and terms & condition regarding Quotation:-

- Interested vendors should submit the quotation as per given format only. Vendors need to specify Manufacturer & brand in the quotation and technical documents and the same supply needs to be supplied.
- Quotations must be submitted in a sealed envelope only. Technical Documents and Financial Bid need to be submitted in separate envelope (Two Envelope Bid). Both Technical and Financial Bid in single envelope will be rejected.
- Both the envelopes should be addressed to - The Dean, Sir J J Group of Hospitals, Mumbai and marked Kind attention to- Surgical store.
- Vendors must write quotation reference no & Last date of submission of quotation on the envelope.
- Any amendments regarding the quotation will be published on website www.ggmccjh.in. Vendors will not be communicated separately regarding the amendments. However if the vendor fails to check any of these amendments on the website then it will be presumed that the vendor has quoted his/ her rates by taking the note of these amendments.
- Rate should be quoted inclusive of all taxes, GST, etc.
- Rate must be written in both figures & words. Rates should be valid for six months from the date of opening the quotation.
- Rate must be quoted for Surgical Items & Instrument complying with the standards for Surgical Items & Instruments (ISO/ISI, etc. whichever is applicable)
- Delivery period is 24 hrs to 03 (Three) days from the receipt of order, as per vitality of Surgical items.
- Analysis test reports (Inhouse and NABL) need to be submitted at the time of goods supply,
- For goods supplied under MJPJAY Other schemes , bills will be passed only after the utilization certificate is received.
- Successful vendors if fails to supply the goods within stipulated delivery period he is liable for the further necessary action, which may deem fit.
- Frequent defaulter bidder will be debarred from participating in the future tenders /quotations called by this office
- Demo Sample will be required to be submitted to Surgical Store, Sir J J Group of Hospitals, Mumbai for all quoted items whenever communicated. The submitted material will not be returned back to the vendor irrespective of qualification or disqualification
- Following documents need to be submitted along with quotation- (To be submitted in Separate Envelope)
 - Bidders FDA, wholesale stocking & Selling Licence
 - Medical Device Licence for quoted Items (clearly highlighting the item quoted)
 - Manufacturers Authorization Letter is mandatory for all above mentioned products.
 - Non conviction certificate from FDA
 - GST registration copy

- Bidder details as per Annexure A
- Affidavit stating that all previous supplies during past 12 months were made as per orders issued by the Dean, Sir, J.J. hospital, Mumbai/any state Government/Public Health Department/ DMER for major supplies.
- Adata (अदत्त) Registration Number at JJ Hospital Mumbai
- WHO GMP certificate of manufacturer, ISO/ISI/CE/EN/ASTM Product Standard Certificate whichever is applicable

NOTE : Last date of submission of quotation: 09 / 11 /2026 before 5.00 pm



**Dean
Sir J J Group of Hospitals,
Mumbai**

Annexure – A

(To be submitted on Bidder's Letterhead, Incomplete Annexure is liable for Rejection)

1. Name and address of the firm: -
2. Registered Head Office Postal address: -
3. Telephone No.
4. E-Mail ID : -
5. Ownership status of the firm- (Maharashtra Govt. / Central Govt./Jt. Sector /co - operative /SSI /Private)
6. Whether bidding as a manufacturer / importer / Authorized Distributor
7. Name of the person & Phone no. who should be contacted by this office in case of emergency.
8. Payee (अदात) Registration Number at Sir J J Hospital, Mumbai.
9. Bank Details: -
 - 1) Bank A/C No. _____
 - 2) IFSC Code: - _____
 - 3) Branch Name & Address: _____
 - 4) Cancelled Cheque: _____
10. PAN number _____
11. GST registration number _____

I / we hereby declare that particulars furnished above are true to the best of my /our

knowledge and belief and that if any of the particulars is found to be materially incorrect / misleading, My /Our quotation shall be rejected. I / we accept all term & condition, also I / we are liable for penal action as per terms specified in the “ terms and conditions of quotation”.

Date: -

Signature of the bidder with official seal and address

ANNEXURE B

Technical Compliance Chart for Quoted Items

Sr No	Name of Items	Specification	Mfg Name	Generic Name/ Brand Name	Medical Device Lic No.	MD Lic Page No.
1	Anatomical Mask	an inflatable cuff for comfort, a standard 22 mm connector, and a disposable or autoclavable (up to 137°C) design, typically made from non-toxic silicone or PVC. 0,1,2,3,4				
2	CPAP Mask	1. The mask should have headgear with a provision to support on the patient face and reduce excess pressure on the patient's nose bridge. 2. Should have provision to maintain 12mm gap between forehead and mask nose bridge frame. 3. The headgear should have an option of providing a personalized fit to help increase comfort and tolerance for patients. 4. Headgear should help in easy sizing fitting and readjusting. 5. Mask should have auto adjusting seal and come with a sizing guide to choose right size for patients. 6. Should have provision for slot for Oro-Naso-Gastric Tubes.				
3	Ultrasonography Jelly	Should have neutral pH, be easy to spread, formaldehyde and salt free, nongreasy, proper adhesion for electrodes, easy squeeze leakproof bottle 250ml quantity, hypoallergenic, nontoxic odourless, high viscosity, high conductivity, water soluble, non staining easy to clean, non irritant to skin. Propylene glycol, Glycerine and phenoxyethanol containing gel Volume 250ml in compressible dispenser				
4	ECG Jelly	Should have neutral pH, be easy to spread, formaldehyde and salt free, nongreasy, proper adhesion for electrodes, easy squeeze leakproof bottle 250ml quantity, hypoallergenic, nontoxic odourless, high viscosity, high conductivity, water soluble, non staining easy to clean, non irritant to skin. Propylene glycol, Glycerine and phenoxyethanol containing gel Volume 250ml in compressible dispenser				
5	Endotracheal Tube Plain (2-9)	Should be available as a transparent tube with radiopaque line in extra-soft PVC. The distal end of the tube should be bevelled at 37+ - 10. The tube should have clear distance markings A standard male ISO adaptor 15 mm to connect the endotracheal tube to the ventilator The distal end of the tube should be bevelled at 37+ -10 2 to 9 mm ID				
6	Endotracheal Tube Cuffed (2-9)	Should be Transparent with Blue radio-opaque line. Should have Low pressure high volume cuff, reduces risk of creasing on inflation thus minimises pressure on tracheal wall. Printed pilot balloon, Smooth Murphy Eye. Should be able to verify tube position within patient. Misting can be clearly seen to confirm intubation The distal end of the tube should be bevelled at 37+ -10 2 to 9 mm ID				
7	Tracheostomy Tube cuffed (Size 6 - 8.5)	1. The Tracheostomy Tube should be a cuffed tube made of thermosensitive material which softens at body temperature to accommodate individual patient anatomy 2. It should have a soft flexible flange which helps in easy fixation 3. It should be provided with a 15 mm connector and can be rotated in 360 degree direction and facilitate the safe connection to breathing system from any direction 4. It should have a radio opaque blue line allows for clear X-ray visualization 5. It should be supplied with cotton neck tape and is sterile and individually blister packed 6. It should be light in weight and non irritant 7. The soft low pressure cuff should dynamically adjust the pressure at the point of contact during the breathing cycle. 8. It should have a pilot balloon with non return valve which facilitates the cuff inflation & deflation size- 6-8.5 mm				
8	Extension Line 25cm	Small bore , high pressure low volume extension tube , with male luer lock at one end and female luer lock at other end . Made up of Medical grade P.V.C, Latex Free. BP/ISO/ISI /CE STANDARDS 150cm				
9	Gigli Saw Wire	Made of martensitic steel (quenched, magnetic steel) and not stainless steel. Hooks at both ends. Should conform to ISO standard				
10	IV Cannula (Size 16 G)	Disposable, Sterile, Single use, individually transparent packing. Colour code according to size. Should have injection port, Snap lid on port, Leak proof, Must be leur fitting Adaptable fixation wings. Material polytetrafluoroethylene (PTFE)/ FEP. Should be non thrombogenic, Non Toxic and Non Pyrogenic. The inner needle must have a bevel for minimal puncture trauma				

Sr No	Name of Items	Specification	Mfg Name	Generic Name/ Brand Name	Medical Device Lic No.	MD Lic Page No.
		Blood flow through needle must be adequate to ensure immediate flow Cannula Should be : a. Thin walled b. X - ray contrast line c. Smooth surface BP/ISO/ISI /CE STANDARDS 16g				
11	IV Cannula (Size 18-22G)	Disposable, Sterile, Single use, individually transparent packing Colour code according to size Should have injection port, Snap lid on port, Leak proof, Must be leak fitting Adaptable fixation wings. Material polytetrafluoroethylene (PTFE)/ FEP. Should be non thrombogenic, Non Toxic and Non Pyrogenic. The inner needle must have a bevel for minimal puncture trauma Blood flow through needle must be adequate to ensure immediate flow Cannula Should be : a. Thin walled b. X - ray contrast line, The distal end of the tube should be bevelled at 37+-10 Smooth surface BP/ISO/ISI /CE STANDARDS (Size 18G-22G)				
12	IV Cannula (Size 24 G)	Disposable, Sterile, Single use, individually transparent packing. Colour code according to size. Should have injection port, Snap lid on port, Leak proof, Must be leak fitting Adaptable fixation wings. Material polytetrafluoroethylene (PTFE)/ FEP. Should be non thrombogenic, Non Toxic and Non Pyrogenic. The inner needle must have a bevel for minimal puncture trauma Blood flow through needle must be adequate to ensure immediate flow Cannula Should be : a. Thin walled b. X - ray contrast line The distal end of the tube should be bevelled at 37+-10 c. Smooth surface BP/ISO/ISI /CE STANDARDS 24G				
13	Ambu Bag Neonatal	It should be self-inflatable and must have pop up valve (non-return valve), attachment for oxygen tube & oxygen reservoir. Bag must be made up of medical grade silicon, latex free double layered which retain sensitivity and it must be resistant to rough use. Inlet end of the bag must have separate port for Oxygen supplement. Outer port must be such that re-breathing valve or non-return valve can be attached. Must be supplied with Oxygen reservoir bag. Should have a bag volume of 250ml. It should be autoclaveable. It should be adaptable to all type of face masks. Must be USFDA or European CE approved product				
14	Rubber Makintosh	Should be made of rubber with double colour, high grade polymer (55% grade A); Should be resistant to high temperature, chemical resistance and fair abrasion resistance. Should be water proof. Colour : Green, Blue, Red, Thickness-0.4mm, Width-90cm Roll of 30 mtr I.S.O. certified, Double colour, water proof ISI Mark 4135: 1974				
15	Mucus Extractor	Transparent Tube mucus extractor for neonatal & infants. Sterile Pyrogen Free, Non Toxic The capacity of the Mucus Extractor is 25ml and a spare cap is also provided for the sealing. Soft and kink resistant tubing with frosted surface A traumatic, soft rounded, beveled open tip. Capacity: 25 ml.				
16	Plastic Gloves	Disposable plastic gloves individually packed. Made of medical grade virgin P.V.C. not less than 50 micron thickness, non toxic with excellent grip Length: 7 Inch Free Size Sleeve Length: 295 mm mid-forearm Type: Non Sterile 100 pcs/Pack				
17	Skin Grafting Blade	Stainless steel blade, sharp blade edge should offer smooth shaving of healthy skin tissue, Individually sterile packed, European CE certified, Expiry date should be written.				
18	T-Connector	T-Connector for Nebulization set				
19	T Piece	For use in recovery for the delivery of oxygen. Connects to 15 mm ET tube connector/ laryngeal mask. Fitted with 22 mm male / 15 mm female patient port. 6" flexible corrugated tube oxygen reservoir with oxygen supply port along with cap. Sterile and individually packed				
20	Flexometallic Endotracheal Tube (Size 6.5 to 8.5)	Used to maintain a patient airway and mechanically ventilate patients. Made from non-toxic, non-irritant medical grade pvc. Kink resistance. Tube can be temporarily straightened for ease of suctioning. Smooth polished Murphy eye, distal end and tapered balloon end for a traumatic intubation and extubation Size 6.5 to 8.5 mmID				
21	Disposable Spinal Needle (Size 20 G)	Disposable, Sterile, Single use, Grooved hub with smooth movement, Length 90mm, Individually transparent packing. Color coded as per size, With introducer. BP/ISO/ISI /CE STANDARDS, 20G				
22	Disposable Spinal Needle (Size 21G)	Disposable, Sterile, Single use, Grooved hub with smooth movement, Length 90mm, Individually transparent packing. Color coded as per size, With introducer. BP/ISO/ISI /CE STANDARDS, 21G				

Sr No	Name of Items	Specification	Mfg Name	Generic Name/ Brand Name	Medical Device Lic No.	MD Lic Page No.
23	Disposable Spinal Needle (Size22G)	Disposable, Sterile, Single use, Grooved hub with smooth movement, Length 90mm, Individually transparent packing. Color coded as per size, With introducer. BP/ISO/ISI /CE STANDARDS, 22G				
24	Disposable Spinal Needle (Size23G)	Disposable, Sterile, Single use, Grooved hub with smooth movement, Length 90mm, Individually transparent packing. Color coded as per size, With introducer. BP/ISO/ISI /CE STANDARDS, 23G				
25	Disposable Spinal Needle (Size24G)	Disposable, Sterile, Single use, Grooved hub with smooth movement, Length 90mm, Individually transparent packing. Color coded as per size, With introducer. BP/ISO/ISI /CE STANDARDS, 24G				
26	Disposable Spinal Needle (Size25G)	Disposable, Sterile, Single use, Grooved hub with smooth movement, Length 90mm, Individually transparent packing. Color coded as per size, With introducer. BP/ISO/ISI /CE STANDARDS, 25G				
27	Negative Wound Suction Set (Romovac Drain) Size 10-18G	Negative Suction Wound Drainage vacuum set should have graduated, bellow of 600 ml capacity, should have 2 radio opaque lined drain catheter Size-10-18G respectively for No.10-18G Should be double packed sterile				
28	N-95 Mask	Should have a shape that should not collapse easily. High filtration efficiency, Good breathability, Tight fit. Quality compliant with standards of surgical N-95 Mask, Fluid resistance at minimum 80mmHg pressure based on ASTM F1862, Quality compliant with standard for particulate respirator that can be worn with full face shield. They should be tested for fluid resistance, filtration efficiency (particulate filtration efficiency and bacterial filtration efficiency), flammability and biocompatibility. NIOSH certified, EN 149:2001 European standard, ISO 22609 standard or equivalent.				
29	Triple Layer Face Mask	Should be triple layer face mask with rust proof nose clip Tie on Mask of Non-woven, hypoallergenic 3 ply construction with filter in between offering > 99% bacterial filtration efficiency (BFE) at 5 micron capacity. Standard size with 4 tie strings and elastic strings				
30	Pulse Oximeter	Pulse Oximeter				
31	Fistula Needle Set 17G	Should be ultra-siliconized thin walled sharp needle with beveled back eye for easy insertion and minimize trauma to the patient. Should be soft and kink resistant medical grade PVC tubing, available with color coded fixed wings, flexible butterfly wings for better grip, needle with back-eye, non toxic, non pyrogenic, sterile. Needle length - 30cms				
32	Sterile Plastic Drape	Sterile disposable poly plain Sheet 100cm x 150cm				
33	Sterile Surgical Gown	SFSMS Polypropylene nonwoven (Non laminated) breathable fabric with antistatic and alcohol repellent treatment, Raglan Sleeves, Adjustable Hook & Loop Binding, Fluorine Chemical Free. CE / ISO				
34	Shoe Cover	Made of Nonwoven material which is water and alcohol repellent. The pair should be of standard size, should be long upto below knee.				
35	Vacuum Blood Collection Tube- EDTA	Lavender double cap with spray dried 5.4mg K2EDTA 13X75mm 3ml draw volume evacuated tube compatible and listed in approved tube list of automated hematology machines kit insert if applicable. Safe for Incineration: Environmental safety data must be provided. Material: tube PET (Poly ethylene terephthalate- virgin material). Holding capacity : 3ml or 4ml. Use: for blood Specimen collection. Standard Dimensions in Millimeters: Vacuum tube 13 mm (D) x 75mm (L). Error margin of biggest vacuum should be 10%, Shelf Life Not less than 12 months, Sterilization Sterile (Gamma radiation), Single Use. CE / ISO				
36	Yankauer Suction Set	Suction Cannula with Suction Tubing				
37	Sterile Cord Clamp	Disposable Umbilical Cord Clamp are made from Non toxic Medical grade polymer. 1. Material: PVC, PE, PP, ABS etc. 2. Sterilize by ethylene oxide gas 3. Supplied sterile in individual peel-able polybag. Or Equivalent				